

Calibration certificate for high temperatures



The loggers that can be provided on request or are always provided with a calibration certificate for temperature are:

- SterilDisk series
- Micro series
- MicroW series (some versions with particular certificates)
- S-Disk J series (some versions with particular certificates)
- Radio series

Standard calibration points are:

- 25°C
- 50°C
- 75°C
- 100°C (last point for Pasteur- and P- series)
- 125°C
- 140°C

