

# Calibration certificate for high temperatures



The loggers that can be provided on request or are always provided with a calibration certificate for temperature are:

- SterilDisk series
- Micro series
- MicroW series (some versions with particular certificates)
- S-Disk J series (some versions with particular certificates)
- Radio series

Standard calibration points are:

- 25°C
- 50°C
- 75°C
- 100°C (last point for Pasteur- and P- series)
- 125°C
- 140°C

Temperatura di riferimento / Reference temperature	Temperatura calibrata / Calibrated temperature	Tolleranza di lettura / Reading tolerance
25.00	24.75	0.25
50.00	49.50	0.50
75.00	74.50	0.50
100.00	99.50	0.50
125.00	124.50	0.50
140.00	139.50	0.50

