

Calibration certificate of pressure



The loggers that are always provided with a calibration certificate for temperature and/or pressure are:

- PressureDisk

Standard calibration points are:

- 25°C
- 50°C
- 75°C
- 100°C
- 125°C
- 140°C
- 10 mbar
- 30 mbar (PressureDisk V only)
- 100 mbar (PressureDisk V only)
- 300 mbar (PressureDisk V only)
- 1 bar (not for PressureDisk V)
- 2 bar (not for PressureDisk V)
- 3 bar (not for PressureDisk V)
- 4 bar (not for PressureDisk V)

Certificato di verifica calibrations / Calibration certificate

Indirizzo / Address: **TECNOSOFT S.p.A.**
 Via Galvani, 4 - 20068, Peschiera Borromeo (MI), Italy
 Tel: +39 02 2692 2888 - Fax: +39 02 2692 2875
 Email: info@tecnosoft.eu - Web: www.tecnosoft.eu

Modello / Model: **PressureDisk**
 Numero di serie / Serial number: **100 000001**

Temperature / Temperature

Temperatura di riferimento / Reference temperature	Temperatura calibrata / Calibrated temperature	Scostamento da riferimento / Error
°C	°C	°C
25.0	25.0	0.00
50.0	50.0	0.00
75.0	75.0	0.00
100.0	100.0	0.00
125.0	125.0	0.00
140.0	140.0	0.00

Segue informazioni / Informazioni Serial: **Modello / Model: PressureDisk V**
 Seriale / Serial: **100 000001**
 Data di produzione / Production date: **10/01/2011**

Accertamento di calibrations / Calibration system accuracy: **0.1 °C**

Pressure / Pressure

Pressione di riferimento / Reference pressure	Pressione calibrata / Calibrated pressure	Scostamento da riferimento / Error
mbar	mbar	mbar
10	10	0.00
30	30	0.00
100	100	0.00
300	300	0.00

Camera di calibrations / Calibration chamber: **Modello / Model: MPF 200**
 Seriale / Serial: **100 000001**
 Data di produzione / Production date: **10/01/2011**

