

Calibration certificate of pressure



The loggers that are always provided with a calibration certificate for temperature and/or pressure are:

- PressureDisk

Standard calibration points are:

- 25°C
- 50°C
- 75°C
- 100°C
- 125°C
- 140°C
- 10 mbar
- 30 mbar (PressureDisk V only)
- 100 mbar (PressureDisk V only)
- 300 mbar (PressureDisk V only)
- 1 bar (not for PressureDisk V)
- 2 bar (not for PressureDisk V)
- 3 bar (not for PressureDisk V)
- 4 bar (not for PressureDisk V)

Certificato di verifica calibrations / Calibration certificate

Modello / Model: PressureDisk
 Numero di serie / Serial number: 100 000001

Temperature / Temperature

Temperatura di riferimento / Reference temperature	Temperatura Calibrata / Calibrated temperature	Scostamento da riferimento / Error
25.00	25.00	0.00
50.00	50.00	0.00
75.00	75.00	0.00
100.00	100.00	0.00
125.00	125.00	0.00
140.00	140.00	0.00

Pressione / Pressure

Pressione di riferimento / Reference pressure	Temperatura / Temperature	Pressione calibrata / Calibrated pressure	Scostamento da riferimento / Error
10.00	25.00	10.00	0.00
30.00	25.00	30.00	0.00
100.00	25.00	100.00	0.00
300.00	25.00	300.00	0.00

Camera di calibrazione / Calibration chamber: 100°C

Camera di calibrazione / Calibration chamber: 100°C

Modello / Model: PressureDisk
 Numero di serie / Serial number: 100 000001



Tecnosoft srl
 Via Galvani, 4, 20068, Peschiera Borromeo (MI), Italy
 T: (+39) 02 2692 2888 - F: (+39) 02 2692 2875
 email: info@tecnosoft.eu - web: www.tecnosoft.eu
 UNI EN ISO 9001:2008 Certiquality/IQNet N. 17733

