

Calibration certificate of humidity



The loggers that can be provided on request or are always provided with a calibration certificate for humidity are:

- HumiStick
- HumiUSB series
- HumiPDF
- RHTPUSB
- PressPDF
- RHTemp series

Standard calibration points, not for RHTemp series, are:

- 5°C
- 30°C
- 50°C
- 30% @ 25°C
- 50% @ 25°C
- 70% @ 25°C

Standard calibration points, for RHTemp series, are:

- 25°C
- 50°C
- 80°C
- 125°C (RHTemp 125 only)
- 30% @ 25°C
- 50% @ 25°C
- 70% @ 25°C

Certificato di verifica calibratore
Certificate of calibration verification

Strumento / Instrument: Data Logger di temperatura e umidità / Temperature and Humidity Data Logger
Costruttore / Manufacturer: Tecnosoft
Modello / Model: HumiStick
Numero di serie / Serial Number: 1000028000740

Temperatura di riferimento / Reference temperature °C	Temperatura misurata / Temperature measured °C	Scostamento rispetto al riferimento / Error
0.00	0.00	0.00
50.00	50.00	0.00
125.00	125.00	0.00

Umidità nominale / Nominale humidity	Umidità registrata / Measured humidity	Scostamento rispetto al riferimento / Error
30%	30%	0.00
50%	50%	0.00
70%	70%	0.00

Stampato in data / Printed on: 01/08/2018
Data di calibratura / Calibration date: 01/08/2018 YAMA / YAMA: 01/08/2011
Camera climatica / Climate chamber: Climatic Chamber Model: MIP 240-20-10002

